

## 血细胞分析仪标准物质

国家编号	样品编号	浓度值 10 <sup>9</sup> /L	SD10 <sup>9</sup> /L	CV/%	规格	单价 元/瓶
GBW(E)090026 血小板标准物质 均值粒径：2.2μm	P001	188.820	1.410	0.75	1.5mL/瓶	60
	P002	114.990	0.46	0.4	1.5mL/瓶	60
	P003	57.56	0.410	0.71	1.5mL/瓶	60
	P102	154.560	2.88	1.86	1.5mL/瓶	60
	P104	77.760	1.19	1.53	1.5mL/瓶	60
	P103 (还有六支)	107.020	2.13	1.99	1.5mL/瓶	60
	P004	312.29	2.93	0.94	1.5mL/瓶	60
	P101	209.900	2.33	1.11	1.5mL/瓶	60
	P005	193.200	4.59	2.38	1.5mL/瓶	60
	P006	58.200	1.76	3.02	1.5mL/瓶	60
GBW(E)090025 白细胞标准物质 均值粒径：5.3μm	W001	2.023	0.48	2.38	1.5mL/瓶	60
	W003	4.017	0.75	1.87	1.5mL/瓶	60
	W004	4.142	0.046	1.11	1.5mL/瓶	60
	W007	6.004	1.12	1.86	1.5mL/瓶	60
	W008	8.078	2.01	2.51	1.5mL/瓶	60
	W009	10.546	2.44	2.31	1.5mL/瓶	60
	W104	7.18	0.15	2.09	1.5mL/瓶	60
	W105	11.70	0.20	1.71	1.5mL/瓶	60
	W106	8.21	0.12	1.46	1.5mL/瓶	60
	W107	7.00	0.16	2.29	1.5mL/瓶	60
	W108	4.98	0.11	2.21	1.5mL/瓶	60
	W109	14.34	0.21	1.46	1.5mL/瓶	60
W1201	10.05	0.23	2.29	1.5mL/瓶	60	

注：SD—粒径分布的标准偏差 CV—粒径分布的变异系数